



05-2025



05-2020

LOT

POB31194

Conforms to EN455 - 1/2

EN374 - 1/2

EN420

AQL = 1.5

In accordance with 93/42/EEC
& 89/686/EEC

合格证

品名称: 一次性PVC
品成分: 聚氯乙烯
规格型号: 医用手套
产日期: 2020-06
产批号: 20000024
检验员: 王
生产日期: 2020-06
规格型号: Q854795-2
产日期: 2020-06
产批号: 20000024
检验员: 王

合格证

品名称: 一次性PVC
品成分: 聚氯乙烯
规格型号: 医用手套
产日期: 2020-06
产批号: 20000024
检验员: 王
生产日期: 2020-06
规格型号: Q854795-2
产日期: 2020-06
产批号: 20000024
检验员: 王

03

GL421303



www.global-consortium

Global Consortium Lin
Charwood House, RH13 9R
Order line: 0800 046

This product is made
the best materials using the
manufacturing techniques, and
the guarantee of qu

GLOBAL CONSORTIUM

vinyl

GL421303



L

vinyl

premium examination gloves

premium exam
powder free



10 X 100 PCS
GL421302

cura[®]
premium vinyl
examination gloves
powder free

Latex free
Single use
Non-sterile
Low protein

Beaded cuff
Textured finish
Conforms to EN455



GL42 Series Cura Vinyl Gloves

Clear vinyl powder free disposable glove



Sectors:

- Healthcare • Food Manufacturing and Preparation • Janitorial • Hair and Beauty

Features

Manufactured from: PVC

Latex free: reducing the risk of hypersensitivity and allergic reactions

Powder Free: Reduces the risk of product contamination

Smooth finish

Code

GL421301	-	Small
GL421302	-	Medium
GL421303	-	Large

Colour

Clear

Sizes

S-XL

Size

Inner Barcode

Outer Barcode

S/7	GL421301	GL421301
M/8	GL421302	GL421302
L/9	GL421303	GL421303
XL/10	GL421304	GL421304

Standards

Manufactured in accordance with Personal Protective Equipment (PPE) by the European PPE Regulation (EU) 2016/425

AQL 1.5 medical grade means greater quality of barrier protection for the wearer

Comply with EN455 to ensure an adequate level of protection

Manufactured to ISO 9001 and ISO 13485 quality standards

Length mm: 230 min

Finger Thickness mm: 0.08 (+/- 0.02)

Palm Thickness mm: 0.08 (+/- 0.02)